

Quality Control Certificate



Estrone ELISA


REF DB52051

LOT 260340

2027-07-16
YYYY-MM-DD

Final Release Results:

	Concentration pg/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	2.842	> 1.100	100%
CAL B	20	2.502	-	88%
CAL C	60	2.082	-	73%
CAL D	200	1.434	-	50%
CAL E	600	0.807	-	28%
CAL F	2000	0.413	-	15%



	Result	Acceptable Range
Std. F % BINDING (OD/ODmax)	15%	< 20%

	Concentration found pg/mL	Target Value pg/mL	Acceptable range lower limit pg/mL	upper limit pg/mL
KIT CONTROL Low CS3827	83.3	86.8	52.1	121.5
KIT CONTROL High CS3827	477.5	493.4	370.1	616.8

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-02-12

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl