

Quality Control Certificate




Androstenedione ELISA

REF DB52161	LOT 260320	2027-02-28 YYYY-MM-DD
--------------------	-------------------	---------------------------------

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	2.633	> 1.000	100%
CAL B	0.1	1.949	-	74%
CAL C	0.3	1.448	-	55%
CAL D	1	0.839	-	32%
CAL E	3	0.454	-	17%
CAL F	10	0.213	-	8%



	Result	Acceptable Range
OD Standard F/% Binding	0.213/8%	< 14%

	Concentration found ng/mL	Target Value ng/mL	Acceptable range lower limit ng/mL	upper limit ng/mL
Control Low CS3674	0.22	0.20	0.12	0.28
Control High CS3674	1.95	2.09	1.57	2.61

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-03-03

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11

E-MAIL: ibl@tecan.com

WEB: www.tecan.com/ibl