

# Quality Control Certificate



## IgA Saliva ELISA

REF DM59171


LOT 6430B

2026-08-31  
YYYY-MM-DD

### Final Release Results:

	Calibrator concentration µg/mL	Measured value OD	Range OD	OD/ODmax
CAL 0	0	0.077	-	3%
CAL 1	6.9	0.263	> 0.105	9%
CAL 2	62	1.131	> 0.450	38%
CAL 3	132	1.792	> 0.715	61%
CAL 4	400	2.953	> 1.180	100%

	Concentration found µg/mL	Acceptable range		
		lower limit µg/mL	target µg/mL	upper limit µg/mL
CONTROL	78.3	53.7	82.7	111.6



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-07-23

**Accepted**

This document has been produced electronically and is valid without a signature.



IBL International GmbH  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)