

Quality Control Certificate



CIC-C1q ELISA

REF DM59211	LOT 6400A	2026-06-30 YYYY-MM-DD
--------------------	------------------	---------------------------------

Final Release Results:

	Concentration µgEq/mL	Result OD	Acceptable range OD
CAL A	0	0.040	< 0.250
CAL B	16	0.651	> 0.260
CAL C	64	2.394	> 0.960

	Result µgEq/mL	Acceptable range	
		Lower Limit µgEq/mL	Upper Limit µgEq/mL
CONTROL -	6.4	3.5	7.5
CONTROL +	47.1	31.5	62.5

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-05-09

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl