

Quality Control Certificate



CIC-C1q ELISA

REF DM59211

LOT 6468A

 2026-10-31
YYYY-MM-DD

Final Release Results:

	Concentration µgEq/mL	Result OD	Acceptable range OD
CAL A	0	0.039	< 0.250
CAL B	16	1.068	> 0.425
CAL C	64	2.683	> 1.075

	Result µgEq/mL	Acceptable range	
		Lower Limit µgEq/mL	Upper Limit µgEq/mL
CONTROL -	5.9	3.3	7.3
CONTROL +	44.2	33.7	66.8

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-09-19

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl