

Quality Control Certificate



CIC-C1q ELISA

REF DM59211

LOT 6580B

 2027-06-30
YYYY-MM-DD

Final Release Results:

	Concentration µgEq/mL	Result OD	Acceptable range OD
CAL A	0	0.058	< 0.250
CAL B	16	0.935	> 0.375
CAL C	64	2.381	> 0.950

	Result µgEq/mL	Acceptable range	
		Lower Limit µgEq/mL	Upper Limit µgEq/mL
CONTROL -	6.7	3.8	8.4
CONTROL +	42.0	28.7	56.9

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-05-13

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl