

Quality Control Certificate



CIC-C1q ELISA

REF **DM59211**

LOT **6621A**

2027-08-31
YYYY-MM-DD

Final Release Results:

	Concentration µgEq/mL	Result OD	Acceptable range OD
CAL A	0	0.080	< 0.250
CAL B	16	0.947	> 0.380
CAL C	64	2.430	> 0.970

	Result µgEq/mL	Acceptable range	
		Lower Limit µgEq/mL	Upper Limit µgEq/mL
CONTROL -	5.6	3.3	7.4
CONTROL +	44.5	29.5	58.5

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-07-01

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl