

Quality Control Certificate



Leptin ELISA

REF MD53001


LOT 040725

2027-02-26
YYYY-MM-DD

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	1	0.111	-	4%
CAL B	10	0.907	-	29%
CAL C	25	1.703	-	55%
CAL D	50	2.353	-	75%
CAL E	100	3.117	-	100%

	Concentration found ng/mL	Acceptable range ng/mL
CONTROL 1	11.48	8.04 - 14.92
CONTROL 2	24.42	17.09 - 31.75



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-08-13

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl