

Quality Control Certificate



Leptin sensitive ELISA

REF MD53011


LOT 210524

2026-06-06
YYYY-MM-DD

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0.05	0.033	-	1%
CAL B	0.5	0.351	-	13%
CAL C	1.5	1.033	-	38%
CAL D	3.5	2.082	-	76%
CAL E	5.0	2.754	-	100%

	Concentration found ng/mL	Acceptable range ng/mL
CONTROL 1	6.96	4.87 - 9.05
CONTROL 2	23.23	16.26 - 30.20



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2024-12-03

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, D-22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl