

# Quality Control Certificate



## IGF-1 ELISA

REF MD58011


LOT 090924

2026-09-18  
YYYY-MM-DD

### Final Release Results:

|       | Concentration<br>ng/mL | Measured<br>value<br>OD | Range<br>OD | OD/ODmax |
|-------|------------------------|-------------------------|-------------|----------|
| CAL A | 2                      | 0.072                   | -           | 3%       |
| CAL B | 5                      | 0.263                   | -           | 11%      |
| CAL C | 15                     | 0.976                   | -           | 39%      |
| CAL D | 30                     | 1.777                   | -           | 72%      |
| CAL E | 50                     | 2.480                   | -           | 100%     |

| In-house control samples |                          |                           |
|--------------------------|--------------------------|---------------------------|
|                          | Target<br>value<br>ng/mL | Acceptable range<br>ng/mL |
| CONTROL 1                | 93                       | 75 - 112                  |
| CONTROL 2                | 399                      | 319 - 479                 |



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-06-25

**Accepted**

This document has been produced electronically and is valid without a signature.



IBL International GmbH  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)