

Quality Control Certificate



IGF-1 ELISA

REF MD58011


LOT 120525

2027-06-11
YYYY-MM-DD

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	2	0.099	-	4%
CAL B	5	0.279	-	13%
CAL C	15	0.914	-	41%
CAL D	30	1.675	-	76%
CAL E	50	2.216	-	100%

In-house control samples		
	Target value ng/mL	Acceptable range ng/mL
CONTROL 1	93	75 - 112
CONTROL 2	428	343-514



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-02-03

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl