

# Quality Control Certificate



## IGFBP-3 ELISA

REF MD58031


LOT 240924

2026-10-23  
YYYY-MM-DD

### Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0.4	0.073	-	3%
CAL B	2	0.304	-	14%
CAL C	6	0.797	-	36%
CAL D	15	1.504	-	67%
CAL E	30	2.240	-	100%

	Target value ng/mL	Acceptable range ng/mL
CONTROL 1	2247	1798-2697
CONTROL 2	4260	3408-5113



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2024-11-20

# Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)