

Quality Control Certificate



Androstenedione RIA (CT)

REF MG11091

LOT 262705K


2026-09-11
YYYY-MM-DD

Final Release Results:

		Measured value	
Total Activity	TA	51798	cpm
Maximum Binding/TA	B0/TA	28.82	%

	Concentration	Measured value	Range	B/B0
	ng/mL	cpm	cpm	
CAL 0	0.0	14929	-	100%
CAL 1	0.11	12240	-	82%
CAL 2	0.31	9805	-	66%
CAL 3	1.04	7244	-	49%
CAL 4	2.84	3678	-	25%
CAL 5	10.0	1457	-	10%

	Measured value	Acceptable range		
	ng/mL	lower limit	target value	upper limit
	ng/mL	ng/mL	ng/mL	ng/mL
CONTROL 1	0.68	0.44	0.65	0.86
CONTROL 2	2.49	1.73	2.47	3.21



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-07-08

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl