

Quality Control Certificate



17-OH-Progesterone RIA (CT)

REF MG12181

LOT 263106AK


2026-10-09
YYYY-MM-DD

Final Release Results:

		Measured value	
Total Activity	TA	80371	cpm
Maximum Binding/TA	B0/TA	34.09	%

	Concentration	Measured value	Range	B/B0
	ng/mL	cpm	cpm	
CAL 0	0	27396	-	100%
CAL 1	0.22	20631	-	75%
CAL 2	0.49	14083	-	51%
CAL 3	1.53	6348	-	23%
CAL 4	5.03	1979	-	7%
CAL 5	16.3	877	-	3%

	Measured value	Acceptable range		
	ng/mL	lower limit	target value	upper limit
	ng/mL	ng/mL	ng/mL	ng/mL
CONTROL 1	1.00	0.67	1.04	1.41
CONTROL 2	2.27	1.61	2.52	3.43



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-08-03

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl