

Quality Control Certificate



Aldosterone RIA (CT)


REF MG13051	LOT 261304AK	2026-05-29 YYYY-MM-DD
--------------------	---------------------	---------------------------------

Final Release Results:

		Measured value	
Total Activity	TA	42971	cpm
Maximum Binding/TA	B0/TA	54.70	%

	Concentration	Measured value	Range	B/B0
	pg/mL	cpm	cpm	
CAL 0	0.0	23506	-	100%
CAL 1	26.5	18681	-	79%
CAL 2	52.9	15411	-	66%
CAL 3	100.0	12356	-	53%
CAL 4	305.0	7321	-	31%
CAL 5	659.0	5098	-	22%
CAL 6	1542.0	3507	-	15%

	Measured value	Acceptable range		
		lower limit	target value	upper limit
	pg/mL	pg/mL	pg/mL	pg/mL
CONTROL 1	191.75	135.00	193.00	251.00



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-03-30

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl