

Quality Control Certificate



Aldosterone RIA (CT)


REF MG13051**LOT** 262908K **2026-09-18**
YYYY-MM-DD

Final Release Results:

		Measured value	
Total Activity	TA	43593	cpm
Maximum Binding/TA	B0/TA	67.98	%

	Concentration	Measured value	Range	B/B0
	pg/mL	cpm	cpm	
CAL 0	0.0	29634	-	100%
CAL 1	26.5	24158	-	82%
CAL 2	52.9	20916	-	71%
CAL 3	100.0	17786	-	60%
CAL 4	305.0	11193	-	38%
CAL 5	659.0	7982	-	27%
CAL 6	1542.0	5112	-	17%

	Measured value	Acceptable range		
	pg/mL	lower limit	target value	upper limit
	pg/mL	pg/mL	pg/mL	pg/mL
CONTROL 1	190.10	135.00	193.00	251.00



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-07-20

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl