Quality Control Certificate

Aldosterone RIA (CT)


REF	LOT	
MG13051	264111K	2026-12-11 YYYY-MM-DD

Final Release Results:

		Measured value	
Total Activity	TA	41921	cpm
Maximum Binding/TA	B0/TA	62.37	%

	Concentration pg/mL	Measured value cpm	Range cpm	B/B0
CAL 0	0.0	26146	-	100%
CAL 1	29.6	20975	-	80%
CAL 2	58.9	18707	-	72%
CAL 3	107.0	15118	-	58%
CAL 4	312.0	10245	-	39%
CAL 5	611.0	7297	-	28%
CAL 6	1532.0	4623	-	18%

	Measured value pg/mL	Acceptable range lower limit pg/mL	target value pg/mL	upper limit pg/mL
CONTROL 1	200.80	135.00	193.00	251.00



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-10-12

This document has been produced electronically and is valid without a signature.

IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl