

Quality Control Certificate



Erythropoietin (EPO) ELISA

REF

NM56011

LOT 3878




2027-02-28

YYYY-MM-DD

Final Release Results:

	Concentration mIU/mL	Measured value OD450nm	Measured value OD405nm	OD/ODmax 405nm
CAL A	0	0.028	0.016	1%
CAL B	10.7	0.209	0.075	5%
CAL C	26.1	0.457	0.162	10%
CAL D	51.5	0.954	0.328	20%
CAL E	152	2.427	0.816	50%
CAL F	451		1.624	100%

	Concentration found mIU/mL	Acceptable range		
		lower limit mIU/mL	target mIU/mL	upper limit mIU/mL
CONTROL 1	19.3	15.3	19.8	24.2
CONTROL 2	196	147	190	232



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-03-10

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11

E-MAIL: ibl@tecan.com

WEB: www.tecan.com/ibl