

Quality Control Certificate



Erythropoietin (EPO) ELISA


REF**NM56011****LOT 3916****2027-08-31**

YYYY-MM-DD

Final Release Results:

	Concentration mIU/mL	Measured value OD450nm	Measured value OD405nm	OD/ODmax 405nm
CAL A	0	0.018	0.012	1%
CAL B	12.4	0.186	0.065	4%
CAL C	27.0	0.414	0.144	9%
CAL D	50.2	0.756	0.259	17%
CAL E	157	2.277	0.762	49%
CAL F	430		1.563	100%

	Concentration found mIU/mL	Acceptable range		
		lower limit mIU/mL	target mIU/mL	upper limit mIU/mL
CONTROL 1	22.3	17.6	22.0	26.4
CONTROL 2	179	123	174	225



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-09-11

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl