

Quality Control Certificate



cAMP RIA (Plasma)


REF RE11021	LOT RCA449	2026-05-21 YYYY-MM-DD
--------------------	-------------------	---------------------------------

Final Release Results:

		Measured value	
Total Activity	TA	60572	cpm
Non-specific binding / TA	NSB/TA	3.0	%
Maximum Binding/TA	(B0-NSB)/TA	34	%

	Concentration pmol/mL	Measured value cpm	Range cpm	(B-NSB) / (B0- NSB)
B0	0	22359	-	100%
CAL A	0.05	20438	-	91%
CAL B	0.15	18196	-	80%
CAL C	0.5	13896	-	59%
CAL D	1.5	9142	-	36%
CAL E	5	5579	-	18%
CAL F	15	3511	-	8%

	Concentration found pmol/mL	Acceptable range		
		lower limit pmol/mL	target pmol/mL	upper limit pmol/mL
CONTROL P	32	19	30	41



Kit Composition:

REF	LOT		REF	LOT	
KRCA151	LYO CONTROL P	H387 2027-04	KRCA281	CONC ASSAYBUF	H237 2026-10
KRCA011	LYO TRACER	H654 2026-05-21	KRCA271	NSB	H236 2026-10
KRCA031	ANTISERUM	H235 2026-10	KRCA251	TRIETHYLAMINE	H206 2026-09
KRCA111	CAL A-I	G835-N 2028-02	KRCA261	ACETANH	H203 2026-09
KRZZ141	CONC SEPREAG	H391 2027-04			

Valid edition of the instructions for use: V2017-09

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-02-27

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl