

Quality Control Certificate



cAMP RIA (Urine)


REF RE11021	LOT RCA449	2026-05-21 YYYY-MM-DD
--------------------	-------------------	---------------------------------

Final Release Results:

		Measured value	
Total Activity	TA	61244	cpm
Non-specific binding/TA	NSB/TA	2.8	%
Maximum Binding/TA	(B0-NSB)/TA	40	%

	Concentration pmol/mL	Measured value cpm	Range cpm	(B-NSB) / (B0-NSB)
B0	0	26177	-	100%
CAL D	1.5	24129	-	92%
CAL E	5	21260	-	80%
CAL F	15	17618	-	65%
CAL G	50	11878	-	42%
CAL H	150	7506	-	24%
CAL I	450	4417	-	11%

	Concentration found nmol/mL	Acceptable range		
		lower limit nmol/mL	target nmol/mL	upper limit nmol/mL
CONTROL U	2.7	1.8	2.4	3.0



Kit Composition:

REF		LOT	
KRCA171	LYO CONTROL U	H386-F	2027-04
KRCA011	LYO TRACER	H654	2026-05-21
KRCA031	ANTISERUM	H235	2026-10
KRCA111	CAL A-I	G835-N	2028-02
KRZZ141	CONC SEPREG	H391	2027-04

REF		LOT	
KRCA281	CONC ASSAYBUF	H237	2026-10
KRCA271	NSB	H236	2026-10

Valid edition of the instructions for use: V2017-09

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-02-26

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl