

Quality Control Certificate



cAMP RIA (Urine)


REF RE11021	LOT RCA450	2026-07-14 YYYY-MM-DD
--------------------	-------------------	---------------------------------

Final Release Results:

		Measured value	
Total Activity	TA	60531	cpm
Non-specific binding/TA	NSB/TA	3.1	%
Maximum Binding/TA	(B0-NSB)/TA	43	%

	Concentration pmol/mL	Measured value cpm	Range cpm	(B-NSB) / (B0-NSB)
B0	0	27740	-	100%
CAL D	1.5	25412	-	91%
CAL E	5	23102	-	82%
CAL F	15	18904	-	66%
CAL G	50	12930	-	43%
CAL H	150	8337	-	25%
CAL I	450	4926	-	12%

	Concentration found nmol/mL	Acceptable range		
		lower limit nmol/mL	target nmol/mL	upper limit nmol/mL
CONTROL U	2.7	1.8	2.4	3.0



Kit Composition:

REF		LOT	
KRCA171	LYO CONTROL U	H386-G	2027-04
KRCA011	LYO TRACER	H691	2026-07-14
KRCA031	ANTISERUM	H235	2026-10
KRCA111	CAL A-I	G835-O	2028-04
KRZZ141	CONC SEPREG	H391	2027-04

REF		LOT	
KRCA281	CONC ASSAYBUF	H237	2026-10
KRCA271	NSB	H236	2026-10

Valid edition of the instructions for use: V2017-09

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: **S. Geweke**

Validation date: **2026-04-23**

Accepted

This document has been produced electronically and is valid without a signature.