

Quality Control Certificate



cAMP RIA (Plasma)


REF RE11021	LOT RCA451	2026-09-09 YYYY-MM-DD
--------------------	-------------------	---------------------------------

Final Release Results:

		Measured value	
Total Activity	TA	61547	cpm
Non-specific binding / TA	NSB/TA	3.1	%
Maximum Binding/TA	(B0-NSB)/TA	25	%

	Concentration pmol/mL	Measured value cpm	Range cpm	(B-NSB) / (B0- NSB)
B0	0	16980	-	100%
CAL A	0.05	15780	-	92%
CAL B	0.15	13818	-	79%
CAL C	0.5	10720	-	58%
CAL D	1.5	7055	-	34%
CAL E	5	4588	-	18%
CAL F	15	3197	-	9%

	Concentration found pmol/mL	Acceptable range		
		lower limit pmol/mL	target pmol/mL	upper limit pmol/mL
CONTROL P	30	19	30	41



Kit Composition:

REF	LOT	2026-09-09	REF	LOT	2026-10-01		
KRCA151	LYO CONTROL P	H387	2027-04	KRCA281	CONC ASSAYBUF	H237	2026-10
KRCA011	LYO TRACER	H749	2026-09-09	KRCA271	NSB	H681	2028-03
KRCA031	ANTISERUM	H680	2028-03	KRCA251	TRIETHYLAMINE	H635	2028-02
KRCA111	CAL A-I	G835-P	2028-06	KRCA261	ACETANH	H636	2028-02
KRZZ141	CONC SEPREAG	H391	2027-04				

Valid edition of the instructions for use: V2017-09

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-06-18

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl