

# Quality Control Certificate



## cAMP RIA (Urine)


<b>REF</b> RE11021	<b>LOT</b> RCA451	<b>2026-09-09</b> YYYY-MM-DD
--------------------	-------------------	---------------------------------

### Final Release Results:

		Measured value	
Total Activity	TA	60933	cpm
Non-specific binding/TA	NSB/TA	2.6	%
Maximum Binding/TA	(B0-NSB)/TA	31	%

	Concentration pmol/mL	Measured value cpm	Range cpm	(B-NSB) / (B0-NSB)
B0	0	20392	-	100%
CAL D	1.5	18509	-	90%
CAL E	5	16716	-	80%
CAL F	15	13635	-	64%
CAL G	50	9417	-	42%
CAL H	150	6342	-	25%
CAL I	450	3933	-	12%

	Concentration found nmol/mL	Acceptable range		
		lower limit nmol/mL	target nmol/mL	upper limit nmol/mL
CONTROL U	2.7	1.8	2.4	3.0



### Kit Composition:

REF		LOT	
KRCA171	LYO CONTROL U	H386-H	2027-04
KRCA011	LYO TRACER	H749	2026-09-09
KRCA031	ANTISERUM	H680	2028-03
KRCA111	CAL A-I	G835-P	2028-06
KRZZ141	CONC SEPREG	H391	2027-04

REF		LOT	
KRCA281	CONC ASSAYBUF	H237	2026-10
KRCA271	NSB	H681	2028-03

Valid edition of the instructions for use: V2017-09

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-06-18

# Accepted

This document has been produced electronically and is valid without a signature.



**IBL International GmbH**  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)