

Quality Control Certificate



cAMP RIA (Urine)

REF RE11021

LOT RCA452


2026-11-05
YYYY-MM-DD

Final Release Results:

		Measured value	
Total Activity	TA	61443	cpm
Non-specific binding/TA	NSB/TA	3.1	%
Maximum Binding/TA	(B0-NSB)/TA	41	%

	Concentration pmol/mL	Measured value cpm	Range cpm	(B-NSB) / (B0-NSB)
B0	0	27353	-	100%
CAL D	1.5	25731	-	94%
CAL E	5	23231	-	84%
CAL F	15	19639	-	70%
CAL G	50	14023	-	48%
CAL H	150	9094	-	28%
CAL I	450	5431	-	14%

	Concentration found nmol/mL	Acceptable range		
		lower limit nmol/mL	target nmol/mL	upper limit nmol/mL
CONTROL U	2.6	1.8	2.4	3.0



Kit Composition:

REF		LOT	
KRCA171	LYO CONTROL U	H386-I	2027-04
KRCA011	LYO TRACER	H807	2026-11-05
KRCA031	ANTISERUM	H680	2028-03
KRCA111	CAL A-I	G835-Q	2028-08
KRZZ141	CONC SEPREAG	H391	2027-04

REF		LOT	
KRCA281	CONC ASSAYBUF	H692	2028-04
KRCA271	NSB	H681	2028-03

Valid edition of the instructions for use: V2017-09

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T. Borchert

Validation date: 2026-08-13

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl

Seite 1/2