

# Quality Control Certificate



## dsDNA-Ab RIA


<b>REF</b> RE19011	<b>LOT</b> RDD383	<b>2026-06-17</b> YYYY-MM-DD
--------------------	-------------------	---------------------------------

**Final Release Results:**

		Measured value	
Total Activity	TA	22402	cpm
Non-specific binding/TA	NSB/TA	3.0	%
Maximum Binding/TA	(Bmax-NSB)/TA	49	%

	Concentration IU/mL	Measured value cpm	Range cpm	(B-NSB) / (Bmax-NSB)
CAL A	0	676	-	0%
CAL B	4.0	1308	-	6%
CAL C	7.0	1660	-	9%
CAL D	14	2767	-	19%
CAL E	40	5907	-	47%
CAL F	80	11756	-	100%

	Concentration found IU/mL	Acceptable range		
		lower limit IU/mL	target value IU/mL	upper limit IU/mL
CONTROL 1	3.3	0.5	3.0	5.5
CONTROL 2	25.1	20.7	25.9	31.1



**Kit Composition:**

REF		LOT	
KRDD151	LYO CONTROL 1	H637-A	2026-09
KRDD161	LYO CONTROL 2	H638-A	2026-09
KRDD011	TRACER	H678	2026-06-17
KRDD111A-F	LYO CAL A-F	H626-A	2026-09
KRDD261	LYO SAMPLEDIL	H568-C	2026-09
KRDD051	AMMSULF	H557	2026-10

Valid edition of the instructions for use: 2022-11

**This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.**

Verified by: **Dr. Spangenberg**

Validation date: **2026-03-18**

Accepted

This document has been produced electronically and is valid without a signature.



**IBL International GmbH**  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)