

Quality Control Certificate



dsDNA-Ab RIA


REF RE19011	LOT RDD384	2026-08-12 YYYY-MM-DD
--------------------	-------------------	---------------------------------

Final Release Results:

		Measured value	
Total Activity	TA	22057	cpm
Non-specific binding/TA	NSB/TA	3.4	%
Maximum Binding/TA	(Bmax-NSB)/TA	52	%

	Concentration IU/mL	Measured value cpm	Range cpm	(B-NSB) / (Bmax-NSB)
CAL A	0	755	-	0%
CAL B	4.5	1469	-	6%
CAL C	10.0	2321	-	14%
CAL D	17	3263	-	22%
CAL E	34	5943	-	45%
CAL F	80	12247	-	100%

	Concentration found IU/mL	Acceptable range		
		lower limit IU/mL	target value IU/mL	upper limit IU/mL
CONTROL 1	2.2	0.5	3.0	5.5
CONTROL 2	23.1	17.7	22.2	26.6



Kit Composition:

REF		LOT	
KRDD151	LYO CONTROL 1	H508-C	2026-11
KRDD161	LYO CONTROL 2	H509-C	2026-11
KRDD011	TRACER	H713	2026-08-12
KRDD111A-F	LYO CAL A-F	H500-C	2026-11
KRDD261	LYO SAMPLEDIL	H657-A	2026-11
KRDD051	AMMSULF	H557	2026-10

Valid edition of the instructions for use: 2022-11

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T. Borchert

Validation date: 2026-05-21

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl