

Quality Control Certificate



dsDNA-Ab RIA

REF RE19011

LOT RDD387


2026-12-02
YYYY-MM-DD

Final Release Results:

		Measured value	
Total Activity	TA	21911	cpm
Non-specific binding/TA	NSB/TA	3.1	%
Maximum Binding/TA	(Bmax-NSB)/TA	52	%

	Concentration IU/mL	Measured value cpm	Range cpm	(B-NSB) / (Bmax-NSB)
CAL A	0	679	-	0%
CAL B	4.0	1079	-	4%
CAL C	7.0	1385	-	6%
CAL D	14	2297	-	14%
CAL E	40	6168	-	49%
CAL F	80	11977	-	100%

	Concentration found IU/mL	Acceptable range		
		lower limit IU/mL	target value IU/mL	upper limit IU/mL
CONTROL 1	3.7	0.1	2.6	5.1
CONTROL 2	25.2	19.3	24.1	28.9



Kit Composition:

REF		LOT	
KRDD151	LYO CONTROL 1	H793-A	2027-03
KRDD161	LYO CONTROL 2	H794-A	2027-03
KRDD011	TRACER	H827	2026-12-02
KRDD111A-F	LYO CAL A-F	H782-A	2027-03
KRDD261	LYO SAMPLEDIL	H808-A	2027-03
KRDD051	AMMSULF	H670	2027-03

Valid edition of the instructions for use:

2022-11

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T. Borchert

Validation date: 2026-09-10

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl