

Quality Control Certificate



Accuracy Controls (ARAb Blocking RIA)

REF	RE21031
-----	---------

Final Release Results:

	Result Inhibition %	Range Inhibition %
ACCURACY CONTROL 1	22	8-27
ACCURACY CONTROL 2	41	22-41
ACCURACY CONTROL 3	52	33-62

The lyophilized ARAb Blocking RIA Accuracy Controls should be dissolved each in 0.5 mL deionized water. Let the controls stand for 15 minutes. Mix by gently swirling. The reconstituted controls can be stored in aliquotes at -20°C until expiry date.

Kit Composition:

REF		LOT	
KRAB173	LYO ACCURACY CONTROL 1	H210-I	2026-10
KRAB183	LYO ACCURACY CONTROL 2	H211-I	2026-10
KRAB193	LYO ACCURACY CONTROL 3	H212-I	2026-10

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T. Borchert

Accepted

Validation date: 2026-04-14

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl