

# Quality Control Certificate



## ARAb Blocking RIA

<b>REF</b> RE21031	<b>LOT</b> RAB281	2026-09-01 YYYY-MM-DD
--------------------	-------------------	--------------------------

### Final Release Results:

			Result		Units
Incubation time			20	3	h
Total Activity (TA)			43142	43180	cpm
Non-specific binding (NSB)			1797	1874	cpm
CONTROL -			8784	10057	cpm
CONTROL +			4450	6421	cpm
NSB/TA			4.2	4.3	% (TA)
CONTROL - / TA	lower limit ≥ 13	upper limit -	20	23	% (TA)
CONTROL +	lower limit 34	upper limit 52	44	46	% Inhibition

### Kit Composition:

<b>REF</b>		<b>LOT</b>	
KRAB013	LYO TRACER	H738A	2026-09-01
KRAB033	NSB REAG	H659-B	2026-12
KRAB153	CONTROL -	H658-B	2026-12
KRAB163	CONTROL +	H660-B	2026-12
KRAB273	CON A SEPH	H567	2026-11
KRAB141	SAMPLEDIL	H519-D	2026-12
KRAR261	BUF	H661-C	2026-09
KRAR253	WASHBUF	H511	2027-08

Valid edition of the instructions for use: V2025-11

<b>This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.</b>	
Verified by:	Dr. Spangenberg
Validation date:	2026-06-10
<b>Accepted</b>	
This document has been produced electronically and is valid without a signature.	

