

# Quality Control Certificate




## 17-beta-Estradiol ELISA

<b>REF</b> RE52041	<b>LOT</b> 42K035	<b>2026-12-25</b> YYYY-MM-DD
--------------------	-------------------	---------------------------------

**Final Release Results:**

	Concentration pg/mL	Measured value OD	Range OD	OD/ODmax
CAL 0	0	1.77	≥1.20	100%
CAL 1	25	1.56	-	88%
CAL 2	100	1.25	-	71%
CAL 3	250	1.09	-	62%
CAL 4	500	0.82	-	46%
CAL 5	1000	0.61	-	34%
CAL 6	2000	0.33	-	19%



	Concentration found pg/mL	Acceptable range		
		lower limit pg/mL	target pg/mL	upper limit pg/mL
CONTROL Low	57.9	34.2	62.2	90.2
CONTROL High	196	128	233	338

**This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.**

Verified by: Dr. Spangenberg

**Accepted**

Validation date: 2025-04-01

This document has been produced electronically and is valid without a signature.



**IBL International GmbH**  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)