

Quality Control Certificate




17-beta-Estradiol ELISA

REF RE52041	LOT 42K065	2027-04-30 YYYY-MM-DD
--------------------	-------------------	---------------------------------

Final Release Results:

	Concentration pg/mL	Measured value OD	Range OD	OD/ODmax
CAL 0	0	2.54	≥1.20	100%
CAL 1	25	2.37	-	93%
CAL 2	100	2.04	-	80%
CAL 3	250	1.72	-	68%
CAL 4	500	1.29	-	51%
CAL 5	1000	0.90	-	35%
CAL 6	2000	0.47	-	19%

	Concentration found pg/mL	Acceptable range		
		lower limit pg/mL	target pg/mL	upper limit pg/mL
CONTROL Low	51.1	34.2	62.2	90.2
CONTROL High	215	128	233	338



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-10-28

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl