

Quality Control Certificate



FSH ELISA

REF RE52121


LOT 15K095

2027-09-30
YYYY-MM-DD

Final Release Results:

	Concentration mIU/mL	Measured value OD	Range OD	OD/ODmax
CAL 0	0	0.02	-	1%
CAL 1	5	0.21	-	9%
CAL 2	10	0.38	-	16%
CAL 3	20	0.79	-	33%
CAL 4	50	1.53	-	65%
CAL 5	100	2.36	≥ 1.20	100%

	Concentration found mIU/mL	Acceptable range		
		lower limit mIU/mL	target mIU/mL	upper limit mIU/mL
Biorad 40461	5.78	2.84	5.17	7.50
Biorad 40462	18.7	10.5	19.1	27.6
Biorad 40463	34.8	18.2	26.5	34.8



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-11-12

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl