

# Quality Control Certificate




## Testosterone ELISA

<b>REF</b> RE52151	<b>LOT</b> ETE136	<b>2026-05-31</b> YYYY-MM-DD
--------------------	-------------------	---------------------------------

**Final Release Results:**

	Concentration ng/mL	manual		
		Measured value OD	Range	OD/ODmax
CAL A	0	2.044	-	100%
CAL B	0.2	1.892	-	93%
CAL C	0.5	1.632	-	80%
CAL D	1	1.348	-	66%
CAL E	2	0.982	-	48%
CAL F	6	0.462	-	23%
CAL G	16	0.183	-	9%



Validation criteria				
	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target value ng/mL	upper limit ng/mL
CONTROL 1	0.75	0.46	0.77	1.08
CONTROL 2	3.93	2.79	3.99	5.19

**Kit Composition:**

REF		LOT	
KETE511	MTP	L428	2026-05
KETE541	ENZCONJ	L430	2026-05
KETE611	CAL A	L416	2026-05
KETE611	CAL B-G	L416	2026-05
KETE651	CONTROL 1	L417	2026-05
KETE661	CONTROL 2	L418	2026-05

REF		LOT	
KIZZZ01	CONC WASHBUF	L345	2027-11
KEZZ991	TMB SUBS	M840-4	2027-02
KEZZ921	TMB STOP	L132	2027-02

Valid edition of the instructions for use: 2024-01

**This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.**

Verified by: S. Geweke

Validation date: 2024-07-24

Accepted

This document has been produced electronically and is valid without a signature.