

Quality Control Certificate



Corticosterone ELISA

REF RE52211


LOT 61K084

2026-06-30
YYYY-MM-DD

Final Release Results:

	Concentration nmol/L	Measured value OD	Range OD	OD/ODmax
CAL 0	0	2.59	≥ 1.20	100%
CAL 1	5	2.15	-	83%
CAL 2	15	1.62	-	63%
CAL 3	30	1.28	-	49%
CAL 4	60	0.97	-	37%
CAL 5	120	0.72	-	28%
CAL 6	240	0.53	-	20%

	Concentration found nmol/L	Acceptable range		
		lower limit nmol/L	target nmol/L	upper limit nmol/L
Control Low	4.45	3.80	5.84	7.88
Control High	60.9	47.3	72.8	98.3



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2024-08-23

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl