

Quality Control Certificate



Progesterone ELISA

REF RE52231


LOT 23K016

2027-01-31
YYYY-MM-DD

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL 0	0.0	2.59	≥ 1.20	100%
CAL 1	0.30	2.01	-	78%
CAL 2	1.25	1.28	-	49%
CAL 3	2.5	0.96	-	37%
CAL 4	5.0	0.73	-	28%
CAL 5	15.0	0.46	-	18%
CAL 6	40.0	0.30	-	12%

	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target ng/mL	upper limit ng/mL
Biorad 40461	0.55	0.41	0.75	1.08
Biorad 40462	9.34	4.94	8.97	13.00
Biorad 40463	28.9	16.8		>40



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-02-04

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl