

Quality Control Certificate




Cortisol Urine ELISA

REF RE52241	LOT 6402A	2026-09-30 YYYY-MM-DD
--------------------	------------------	---------------------------------

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL 0	0	3.261	> 1.305	100%
CAL 1	1	2.356	> 0.940	72%
CAL 2	5	1.424	> 0.570	44%
CAL 3	30	0.615	> 0.245	19%
CAL 4	200	0.205	> 0.080	6%

	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target ng/mL	upper limit ng/mL
CONTROL LOW	3.0	1.9	3.2	4.4
CONTROL HIGH	60.4	47.2	67.5	87.7



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-05-13

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl