

# Quality Control Certificate




## Cortisol Urine ELISA

<b>REF</b> RE52241	<b>LOT</b> 6449C	<b>2026-09-30</b> YYYY-MM-DD
--------------------	------------------	---------------------------------

### Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL 0	0	2.739	> 1.095	100%
CAL 1	1	2.031	> 0.810	74%
CAL 2	5	1.400	> 0.560	51%
CAL 3	30	0.652	> 0.260	24%
CAL 4	200	0.169	> 0.070	6%

	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target ng/mL	upper limit ng/mL
CONTROL LOW	4.4	2.6	4.3	6.0
CONTROL HIGH	62.4	46.7	66.7	86.7



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-12-16

**Accepted**

This document has been produced electronically and is valid without a signature.



IBL International GmbH  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)