

Quality Control Certificate




Aldosterone ELISA

REF RE52301	LOT 89K036	2027-06-30 YYYY-MM-DD
--------------------	-------------------	---------------------------------

Final Release Results:

	Concentration pg/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	1.71	≥ 1.20	100%
CAL B	20	1.27	-	74%
CAL C	80	0.77	-	45%
CAL D	200	0.44	-	26%
CAL E	500	0.24	-	14%
CAL F	1000	0.16	-	9%



	Concentration found pg/mL	Target Value pg/mL	Acceptable range lower limit pg/mL	upper limit pg/mL
Control Low	54.9	50.7	27.9	73.5
Control High	479	542	298	786

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-04-28

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl