

Quality Control Certificate



Aldosterone ELISA


REF RE52301

LOT 89K055

2026-07-31
YYYY-MM-DD

Final Release Results:

	Concentration pg/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	2.53	≥ 1.20	100%
CAL B	20	2.04	-	81%
CAL C	80	1.04	-	41%
CAL D	200	0.53	-	21%
CAL E	500	0.23	-	9%
CAL F	1000	0.15	-	6%



	Concentration found pg/mL	Target Value pg/mL	Acceptable range lower limit pg/mL	upper limit pg/mL
Control Low	35.5	52.7	29.0	76.4
Control High	352	442	243	641

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-05-26

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl