

Quality Control Certificate




Cortisol Saliva ELISA

REF RE52611	LOT ECO168	2026-06-30 YYYY-MM-DD
--------------------	-------------------	---------------------------------

Final Release Results:

	Concentration µg/dL	Measured value OD	manual	
			Range OD	OD/ODmax
CAL A	0	2.062	> 1.000	100%
CAL B	0.015	1.705	-	83%
CAL C	0.04	1.372	-	67%
CAL D	0.17	0.853	-	41%
CAL E	0.70	0.490	-	24%
CAL F	3.00	0.196	-	10%

	Concentration found µg/dL	Acceptable range		
		lower limit µg/dL	target µg/dL	upper limit µg/dL
CONTROL 1	0.077	0.038	0.077	0.115
CONTROL 2	0.56	0.40	0.58	0.75



Kit Composition:

REF		LOT	
KECO651	CONTROL 1	L463A-2A	2026-07
KECO661	CONTROL 2	L464A-2A	2026-07
KECO511	MTP	L781	2026-06
KECO541	ENZCONJ	L804	2027-01
KECO611	CAL A	L461-2A	2026-07
KECO621	CAL B-F	L462-2A	2026-07
KIZZ01	CONC WASHBUF	L780	2028-06

REF		LOT	
KEZZ991	TMB SUBS	L704-2	2027-10
KEZZ921	TMB STOP	L720	2027-11

Valid edition of the instructions for use: 2024-02

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: M.Dries

Validation date: 2025-01-23

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl