

# Quality Control Certificate




## Cortisol Saliva ELISA

<b>REF</b> RE52611	<b>LOT</b> ECO169	<b>2027-01-31</b> YYYY-MM-DD
--------------------	-------------------	---------------------------------

**Final Release Results:**

	Concentration µg/dL	manual		
		Measured value OD	Range OD	OD/ODmax
CAL A	0	2.105	> 1.000	100%
CAL B	0.015	1.868	-	89%
CAL C	0.04	1.496	-	71%
CAL D	0.17	0.936	-	44%
CAL E	0.70	0.568	-	27%
CAL F	3.00	0.214	-	10%



	Concentration found µg/dL	Acceptable range		
		lower limit µg/dL	target µg/dL	upper limit µg/dL
CONTROL 1	0.077	0.039	0.078	0.117
CONTROL 2	0.45	0.36	0.51	0.67

**Kit Composition:**

REF		LOT	
KECO651	CONTROL 1	K184	2027-01
KECO661	CONTROL 2	K185	2027-01
KECO511	MTP	K173	2027-12
KECO541	ENZCONJ	K186	2027-07
KECO611	CAL A	K182	2027-01
KECO621	CAL B-F	K183	2027-01
KIZZ01	CONC WASHBUF	K246	2028-12

REF		LOT	
KEZZ991	TMB SUBS	L919	2028-06
KEZZ921	TMB STOP	K133	2028-05

Valid edition of the instructions for use: 2024-02

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-07-16

**Accepted**

This document has been produced electronically and is valid without a signature.



**IBL International GmbH**  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)