

Quality Control Certificate




Cortisol Saliva ELISA

REF RE52611	LOT ECO170	2027-04-30 YYYY-MM-DD
--------------------	-------------------	---------------------------------

Final Release Results:

	Concentration µg/dL	Measured value OD	manual	
			Range OD	OD/ODmax
CAL A	0	2.408	> 1.000	100%
CAL B	0.015	2.280	-	95%
CAL C	0.04	1.913	-	79%
CAL D	0.17	1.245	-	52%
CAL E	0.70	0.757	-	31%
CAL F	3.00	0.300	-	12%

	Concentration found µg/dL	Acceptable range		
		lower limit µg/dL	target µg/dL	upper limit µg/dL
CONTROL 1	0.083	0.042	0.084	0.126
CONTROL 2	0.59	0.46	0.66	0.86



Kit Composition:

REF		LOT	
KECO651	CONTROL 1	K184A-2A	2027-05
KECO661	CONTROL 2	K185A-2A	2027-05
KECO511	MTP	K382	2027-04
KECO541	ENZCONJ	K399	2027-11
KECO611	CAL A	K182-2A	2027-05
KECO621	CAL B-F	K183A-2A	2027-05
KIZZ01	CONC WASHBUF	K391	2029-04

REF		LOT	
KEZZ991	TMB SUBS	K132-6	2028-09
KEZZ921	TMB STOP	K380	2028-09

Valid edition of the instructions for use: 2024-02

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-12-03

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl