

Quality Control Certificate




Cortisol Saliva ELISA

REF RE52611	LOT ECO171	2027-11-30 YYYY-MM-DD
--------------------	-------------------	---------------------------------

Final Release Results:

	Concentration µg/dL	Measured value OD	manual	
			Range OD	OD/ODmax
CAL A	0	2.120	> 1.000	100%
CAL B	0.015	1.953	-	92%
CAL C	0.04	1.564	-	74%
CAL D	0.17	0.986	-	47%
CAL E	0.70	0.601	-	28%
CAL F	3.00	0.228	-	11%

	Concentration found µg/dL	Acceptable range		
		lower limit µg/dL	target µg/dL	upper limit µg/dL
CONTROL 1	0.080	0.037	0.075	0.112
CONTROL 2	0.57	0.43	0.62	0.80



Kit Composition:

REF		LOT	
KECO651	CONTROL 1	K754	2027-11
KECO661	CONTROL 2	K755	2027-11
KECO511	MTP	K734	2027-11
KECO541	ENZCONJ	K756	2028-05
KECO611	CAL A	K752	2027-11
KECO621	CAL B-F	K753	2027-11
KIZZ01	CONC WASHBUF	K738	2029-10

REF		LOT	
KEZZ991	TMB SUBS	K610	2029-08
KEZZ921	TMB STOP	K595	2029-03

Valid edition of the instructions for use: 2024-02

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T. Borchert

Validation date: 2026-06-02

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl