

Quality Control Certificate



DHEA-S Saliva ELISA

REF RE52661


LOT 6451C

2026-12-31
YYYY-MM-DD

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL 0	0	2.597	>1.040	100%
CAL 1	0.2	2.365	>0.945	91%
CAL 2	1.0	1.556	>0.620	60%
CAL 3	3.0	0.932	>0.375	36%
CAL 4	12.0	0.306	>0.120	12%

	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target ng/mL	upper limit ng/mL
Low Control	1.59	1.15	1.92	2.68
High Control	6.30	4.25	6.26	8.26



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-12-16

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl