

Quality Control Certificate



Estrone Saliva ELISA

REF RE52681


LOT 98K113

2027-08-01
YYYY-MM-DD

Final Release Results:

	Calibrator concentration pg/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	2.42	≥ 1.20	100%
CAL B	3	2.09	-	86%
CAL C	12.3	1.86	-	77%
CAL D	37	1.56	-	64%
CAL E	111	1.16	-	48%
CAL F	333	0.68	-	28%

	Concentration found pg/mL	Acceptable range		
		lower limit pg/mL	target pg/mL	upper limit pg/mL
Control 1	8.27	5.28	9.6	13.9
Control 2	225	153	235	317



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2023-12-21

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl