

Quality Control Certificate



Estrone Saliva ELISA

REF RE52681


LOT 98K125

2026-12-31
YYYY-MM-DD

Final Release Results:

	Calibrator concentration pg/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	1.73	≥ 1.20	100%
CAL B	3	1.66	-	96%
CAL C	12.3	1.48	-	86%
CAL D	37	1.12	-	65%
CAL E	111	0.73	-	42%
CAL F	333	0.44	-	25%

	Concentration found pg/mL	Acceptable range		
		lower limit pg/mL	target pg/mL	upper limit pg/mL
Control 1	11.0	5.08	9.2	13.4
Control 2	170	113	174	235



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-01-08

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl