

Quality Control Certificate



C-Peptide ELISA

REF RE53011


LOT 20K095

2026-09-30
YYYY-MM-DD

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0.0	2.06	≥ 1.20	100%
CAL B	0.2	1.75	-	85%
CAL C	0.7	1.61	-	78%
CAL D	2.0	1.23	-	60%
CAL E	6.0	0.78	-	38%
CAL F	16.0	0.30	-	15%

	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target ng/mL	upper limit ng/mL
Biorad 40462	4.25	2.20	4.00	5.80
Biorad 40463	6.73	3.39	6.16	8.9



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-12-10

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl