

Quality Control Certificate




25-OH-Vitamin D ELISA

REF RE53041	LOT EVD139	2026-08-31 YYYY-MM-DD
--------------------	-------------------	---------------------------------

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	1.989	> 1.000	100%
CAL B	6.00	1.879	-	94%
CAL C	12.0	1.710	-	86%
CAL D	24.0	1.304	-	66%
CAL E	48.0	0.720	-	36%
CAL F	96	0.234	-	12%



	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target value ng/mL	upper limit ng/mL
CONTROL 1	15.7	9.9	14.1	18.4
CONTROL 2	40.6	29.0	41.5	53.9

Kit Composition:

REF		LOT	
30110293	MTP	K782	2029-01
30110294	CAL A	K684	2027-08
30110295	CAL B-F	K685	2027-08
30110296	CONTROL 1	K686	2027-08
30110298	CONTROL 2	K687	2027-08
30110301	ENZCONJ	K785	2027-12

REF		LOT	
30110300	BIOTIN CONC	K784	2027-12
30110299	INCBUF	K783	2027-12
KIZZZ01	WASHBUF CONC	K645	2029-03
KEZZ991	TMB SUBS	K610-2	2029-03
KEZZ921	TMB STOP	K654	2029-03

Valid edition of the instructions for use: 2023-07

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: **Dr. Spangenberg**

Validation date: **2026-05-20**

Accepted

This document has been produced electronically and is valid without a signature.