

Quality Control Certificate



Chromogranin A ELISA

REF RE53071


LOT AA345

2027-09-30
YYYY-MM-DD

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	0.034	-	1%
CAL B	25.6	0.090	-	3%
CAL C	101.0	0.340	-	13%
CAL D	260.5	0.803	-	31%
CAL E	795.8	1.613	-	62%
CAL F	2166.8	2.609	-	100%

	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target ng/mL	upper limit ng/mL
CONTROL 1	34.7	20.0	28.6	37.1
CONTROL 2	147.9	104.0	148.6	193.1




This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-10-15

Accepted

This document has been produced electronically and is valid without a signature.

 IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl